

Oct 15



RF Amplifier

**Including: Basic Modulator Alignment
D1345-aQ110-4**

Instruction Manual RFA0110-1-x Series

Models -

RFA0110-1 : 90-140MHz, amplifier module, 12W output

Options -x:

ISOMET CORP, 5263 Port Royal Rd, Springfield, VA 22151, USA.
Tel: (703) 321 8301, Fax: (703) 321 8546, e-mail: isomet@isomet.com
www.ISOMET.com
ISOMET (UK) Ltd, 18 Llantarnam Park, Cwmbran, Torfaen, NP44 3AX, UK.
Tel: +44 1633-872721, Fax: +44 1633 874678, e-mail: isomet@isomet.co.uk



1. GENERAL


The RFA0110-1 Power Amplifier, figure 1, contains a fixed gain broadband RF amplifier specifically designed to operate with Isomet acousto-optic devices such as the D1345-aQ110. The driver requires low level RF signal from a suitable frequency source such as the Isomet iMS4-L frequency synthesizer. Figure 2 is a functional block diagram of the driver.

The rise and fall response time for the amplifier is approx' 25nsec.

This amplifier is designed to operate at full rated power into a 50Ω load with 100% duty cycle.

Trace 1 = RF output

Trace 2 = Sync signal



Water cooling is optional.

The heatsink temperature must not exceed 70°C.


SERIOUS DAMAGE TO THE AMPLIFIER MAY RESULT IF THE TEMPERATURE EXCEEDS 70°C. SERIOUS DAMAGE TO THE AMPLIFIER MAY ALSO RESULT IF THE RF OUTPUT CONNECTOR IS OPERATED OPEN-CIRCUITED OR SHORT-CIRCUITED.

A low impedance d-c power source is required. The operating voltage is +24V or +28Vdc at a current drain of approximately 3A (4A maximum). The external power source should be regulated to ± 2% and the power supply ripple voltage should be less than 200mV for best results.

Higher RF output power is achieved at 28Vdc.

2.1 LED INDICATOR

The front panel tri-colour LED indicates the operating state.



LED Stack

RED

The top LED will illuminate RED when 24Vdc supply is applied.

Normal condition is ON

YELLOW

The middle LED will illuminate YELLOW, when the RF Gate input is valid.

(Default condition = valid, unless a connection is made to pin7 of the D-type)

Normal condition is ON, but may be OFF if the above conditions are not met

GREEN

The lower LED will illuminate GREEN when the following signals are all true:

- 1) RF DC power is applied and
- 2) Gate signal is valid and
- 3) Amplifier and AO thermal interlocks are valid *.

Normal condition is ON

*** Thermal Interlocks**

The AOM and Driver are fitted with thermostatic switches which will switch open circuit if a predetermined temperature is exceeded. These thermal interlocks will reset once the AO device and / or RF driver are cooled below this temperature.

- The driver thermal switch over-temperature threshold is 50deg C
- The AOD thermal switch over-temperature threshold is 36deg C

The hysteresis of the thermal switches is 7-10deg C.

Once in a fault state the coolant temperature will need to be reduced to reset the thermal switches.



3.0 INSTALLATION AND ADJUSTMENT

Please refer to the Synthesizer manual for frequency, phase and amplitude control of the input signals.

3.1 For water cooled AO devices;

Connect cooling water at a flow of more than 1 litres/minute at < 20 deg.C to both the RF amplifier and AO device. Due to the RF power dissipated in the AO modulator, it is paramount that the device is operated only when water cooling is circulating. For optimum AO performance, ensure the flow rate is greater than 1 litre /minute at < 20 deg.C.

3.2 For conduction cooled AO devices;

Ensure the AOM and RF driver are mounted to a good heat conduction surface

3.3 With no d-c power applied, connect the + 24V (or +28V) DC in to the screw terminal. DO NOT APPLY POWER.

3.4 Connect the RF output BNC jacks to the acousto-optic deflector (or a 50Ω RF load, if to measure the modulator RF output power).

Connection order depends on the orientation as shown on page 12.

Relative phase delay depends on the input source.

3.3a Connect the RF input SMA jacks to the external frequency source outputs (90 – 130MHz, 1mW max, 50Ω, each input).

3.4 Connect the Interlock of the acousto-optic device to the mating connector of the RF driver (Binder 3pin snap connector).

The interlock connection becomes open circuit disabling the RF output, if the temperature of the modulator exceeds 36°C or the internal driver temperature exceeds 50°C. The LED indicator illuminates when the Interlocks are closed and the RF is enabled (see Section 2).

3.5 Adjustment of the RF output power is best done with amplifier connected to the acousto-optic modulator.

- 3.6 **The optimum RF power level required for the modulator to produce maximum first order intensity will be different at various laser wavelengths. Applying RF power in excess of this optimum level will cause a decrease in first order intensity (a false indication of insufficient RF power) and makes accurate Bragg alignment difficult. It is therefore recommended that initial alignment be performed at a low RF power level.**
- 3.7 Set the input power level to give approximately 4W output.
- 3.8 Apply DC power to the amplifier.
- 3.9 Apply a constant RF input to the input SMA connector of the RFA0110-1 (ref: 3.7)

Input the laser beam toward the centre of either aperture of the AO device. Ensure the polarization is vertical with respect to the base and the beam height does not exceed the active aperture height of the AOM/AOD.


Start with the laser beam normal to the input optical face of the AOD and very slowly rotate the AOD (see page 12 for configurations.)

- 3.10 Observe the diffracted first-order output from the acousto-optic modulator and the undeflected zeroth order beam. Adjust the Bragg angle (rotate the modulator) to maximise first order beam intensity.
- 3.11 After Bragg angle has been optimized, slowly increase the RF input power until maximum first order intensity is obtained. This should occur at < 10W peak for the D1345-aQ110-4 at 374nm.

3.12 To equalise deflection efficiency at the extremes of the scan, alternate between the minimum and maximum desired frequencies and adjust Bragg angle to give the same efficiency for both. (Note: the photo detector or light power meter will require repositioning for the two angles.) .

Typical swept frequency response at 374nm

First order sweep



RF power on 25% duty cycle



4. MAINTENANCE

4.1 Cleaning

It is of utmost importance that the optical apertures of the deflector optical head be kept clean and free of contamination. When the device is not in use, the apertures may be protected by a covering of masking tape. When in use, frequently clean the apertures with a pressurized jet of filtered, dry air.

It will probably be necessary in time to wipe the coated window surfaces of atmospherically deposited films. Although the coatings are hard and durable, care must be taken to avoid gouging of the surface and leaving residues. It is suggested that the coatings be wiped with a soft ball of brushed (short fibres removed) cotton, slightly moistened with clean alcohol. Before the alcohol has had time to dry on the surface, wipe again with dry cotton in a smooth, continuous stroke. Examine the surface for residue and, if necessary, repeat the cleaning.

4.2 Troubleshooting

No troubleshooting procedures are proposed other than a check of alignment and operating procedure. If difficulties arise, take note of the symptoms and contact the manufacturer.

4.3 Repairs

In the event of deflector malfunction, discontinue operation and immediately contact the manufacturer or his representative. Due to the high sensitive of tuning procedures and the possible damage which may result, no user repairs are allowed. Evidence that an attempt has been made to open the optical head will void the manufacturer's warranty.

RFA 0110-2 Standard Version

Connection Summary

1.0 'D' Type Control Connection

<u>Signal</u>	<u>Type</u>	<u>Pin out connection</u>
Digital Gate CMOS high (12V logic) or NC = ON CMOS low (0.0v<V<1v) = OFF	Input	Signal pin 7 Return pin 2

2.0 Coaxial SMA (2)

Low level RF Input

Frequency range 90 – 130MHz Typical
 80 – 140MHz Maximum

Power level 0dBm (1mW) Typical
 3dBm (2mW) Maximum

3.0 Interlock connection

AOM Thermal Interlock Plug
(OK = connected contacts
1-2)

RF Driver INT Plug
(OK = connected
contacts 1-2)



The interlock signal must be connected. Contacts closed for normal operation.

4.0 Mounting Holes

4 x M5




Figure 1: Driver Installation




Figure 2: Driver Block Diagram





Figure 4: Typical Connection Configuration

Diagram shows typical beam alignment..


Laser can be input either side of AOM.

See connection options below.

Orientation options



Schematic of a single electrode acousto optic deflector and tunable driver



The input bragg angle, relative to a normal to the optical surface and in the plane of deflection is :

$$\theta_{BRAGG} = \frac{\lambda \cdot fc}{2 \cdot v}$$

The separation angle between the zeroth order and mid scan point of the first order is :

$$\theta_{SEP} = \frac{\lambda \cdot fc}{v}$$

The first order scan angle is :

$$\theta_{SCAN} = \frac{\lambda \cdot \delta f}{v}$$

where:

λ = wavelength

fc = centre frequency = 110MHz

v = acoustic velocity of interaction material = 5.7mm/usec (a-Quartz)

$d = 1/e^2$ beam diameter

Figure 5. Deflection System